

Declaration of conformity

Manufacturer:

Name: Shenzhen Jumper Medical Equipment Co., Ltd

Address: D Building, No. 71, Xintian Road, Fuyong Street, Baoan, Shenzhen, Guangdong, China

European Representative:

Name: Weikang ltd t/a welkang tech consulting

Address: Suite B, 29 Harley Street London W1g 9QR, England United Kingdom.

Product:

Name: Infrared Thermometer

Model: JPD-FR202

Classification:

Infrared thermometer is classified as II a device according to MDD 93/42/EEC annex IX rule 10

Conformity assessment Route:

Annex II.3 is chosen for the conformity certification pathway.

We, the manufacturer, herewith declare that the stated medical devices meet the transposition into national law, the provisions of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices including, at 21 March 2010, the amendments by Council Directive 2007/47/EEC.

The validity period of this declaration of conformity is limited by the issuing of a revised declaration of conformity after change of the product and/or by the expiration date of the related Annex II certificate issued by the notified body.

Standards:


All applicable harmonized Standard (published in the Official Journal of the European Communities)

Notified Body: **MEDCERT GmbH**
Pilatuspool 2, 20355 Hamburg

(EC) Certificate(s): 0482

Place, Date of Issue: *Shenzhen, Dec 16, 2019*

Signature: _____



Crystal Kang